

## ADDITIONAL INFORMATION

**Caution:**

Federal law restricts this device to sale, distribution, and use by or on the order of a physician.

**Warnings:**

Read these instructions for use before using for the first time and familiarise yourself with the operation of the system. Test the bone cement properties and handling of **PALAMIX**<sup>®</sup> in advance under your specific surgical conditions.

**Precautions:**

The **PALAMIX**<sup>®</sup> vacuum mixing system should be examined for damage prior to use. If the sterile packaging is damaged or the shelf life exceeded, the **PALAMIX**<sup>®</sup> vacuum mixing system must be discarded.